



EU Declaration of Conformity – Pulsar Measurement OCF 6.1 Open Channel Flow & Tank Level Meter

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Relevant Directive(s)	2014/30/EU – EMC directive and its amending directives 2014/35/EU – Low Voltage directive and its amending directives 2011/65/EU – RoHS directive and its amending directives
Manufacturer’s name	Greyline Instruments Canada Inc.
Manufacturer’s address	16456 Sixsmith Drive Long Sault, ON K0C 1P0 Canada
Apparatus	Open channel flow & tank level meter including wall-mount electronics, and transducer with built-in cable
Models	OCF 6.1 with any PZ level sensor
Type of equipment	Measurement and process control
Standards applied	EN 61326-1:2013 EMC, equipment class industrial EN 61010-1:2010+A1:2019 Safety requirements for electrical equipment for measurement, control, and laboratory use

I declare that the apparatus named above has been tested and complies with the relevant sections of the above referenced standards & directives.

Signed:

Date: November 16, 2022
Rev. 1.0

Name & function

Ken Elander, Sr Product
Manager

Pulsar Measurement

