



09

EC - DECLARATION OF CONFORMITY

We,

Dwyer Instruments, Inc.
P.O. Box 373
Michigan City, IN 46361, USA
(219) 879-8868

declare under our sole responsibility that our Series DM-1100 Digital Differential Pressure Gage to which this declaration relates are in conformity with the following EC Directives and harmonized standards:

Directive 104/108/EC (EMC)

CISPR11: 2004 Industrial, Scientific and Medical (ISM) Radio-Frequency Equipment – Electromagnetic Disturbance Characteristics – Limits and Methods of Measurement

CENELEC EN 613326-1: 2006 Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements Part - General Requirements

IEC 61000-4-2: 2000 Part 4-2 Test and Measurement Techniques – Electrostatic Discharge Immunity Test

IEC 61000-4-3: 2002 Part 4-3 Testing and Measurement Techniques – Radiated, Radio-Frequency, Electromagnetic Field Immunity Test (Test Level: 10 V/m; 80-1000 MHz, 3 V/m; 1400-2700 MHz and 1 V/m; 2000-2700 MHz)

IEC 61000-4-8: 2000 Part 4-8 Testing and Measurement Techniques – Power Frequency Magnetic Field Immunity Test

The authorized representative located within the Community is:

Dwyer Instruments Ltd
Unit 16, The Wye Estate, London Road
High Wycombe, Bucks HP11 1LH-U.K.
(+44) (0) 1494 461707

On behalf of Dwyer Instruments, Inc.

Senior Project Engineer

Michigan City, Indiana, USA
10/22/09